


1st July 2008



Life Safety Distribution AG
Wilstrasse 11
8610 Uster
Switzerland
Tel +41 (0)44 943 44 38
Fax +41 (0)44 732 69 83
Web www.citytech.com

CRITICAL COMMUNICATION. ACTION REQUIRED!

Product Recall Notice
4CF Carbon Monoxide Sensor
Serial Numbers starting from 10199194 to 12844839

As indicated in our letter of 7th May 2008 regarding the supply of 4CF Sensors, we have continued to investigate the failure rate experienced in these sensors, and have carried out a more extensive sampling process. From this work we have concluded that higher rates of failure are demonstrated since a Design Change in May 2007.

We have therefore decided, as a precautionary measure, to recall 4CF Product with the above serial numbers.

These further investigations also more clearly identified the cause of the problem, and the replacement Sensors have been upgraded to resolve the problem.

We will replace all affected 4CF Sensors free of charge. However, please return to us each 4CF Sensor that is replaced so that we can record the progress of this recall.

AFFECTED SENSORS

All 4CF Sensors with Serial Numbers in the range **10199194 to 12844839** marked Rev. 3 are affected and should be replaced.

Please note that only the 4CF Sensors with the Serial Numbers in the range detailed above are affected. All other 4CF Sensors are unaffected and NO ACTION is necessary for these Sensors.

Please pass to all relevant personnel within your organisation